

Specification

General purpose solid medium containing animal and plant peptone according to Pharmacopoeial Harmonised Method and ISO Standards.

Presentation

10 Prepared bottles
Bottles 250 ml
with: 200 ± 5 ml

Packaging Details

1 box with 10 bottles 250 ml
Non injectable cap

Shelf Life

16 months

Storage

8-25°C

Composition

Composition (g/l):

Casein peptone.....	15.0
Soy peptone.....	5.00
Sodium chloride.....	5.00
Agar.....	15.0

Description /Technique

Description

TSA is a widely used medium containing two peptones which support the growth of a wide variety of organisms, even that of very fastidious ones such as *Neisseria*, *Listeria*, *Brucella*, etc. It is frequently used for routine diagnostic purposes due to its reliability and its easily reproducible results.

The following list includes some of its most common applications:

1. Sensitivity testing, either by the Kirby-Bauer system or by following the WHO guidelines. Both systems recommend the use of the Mueller-Hinton Agar or verification purposes.
2. The medium provides, with added blood, perfectly defined haemolysis zones, while preventing the lysis of erythrocytes due to its sodium chloride content.
3. It can be used for the preparation of an exceptionally nutrient 'chocolate' agar, thanks to the richness of its peptones.
4. In a reducing environment or with a CO₂ enriched atmosphere, it provides an excellent medium for the isolation of *Brucella* and *Neisseria*. It may be made selective by using additives.
5. Most streptococci grow in this medium though clear differences can be observed from one species to another.
6. Tryptic Soy Agar can be used as a selective medium for the count of urine samples although differentiation must be done on selective differential media.
7. Several tests for the differentiation and identification of staphylococci can be performed on this medium, provided suitable additives are used.
8. Yeast, particularly *Candida* species, can grow in this medium forming very characteristic colonies.
9. Chromogenic *Pseudomonas* frequently produce pigmentation on TSA and are therefore easily recognized.
10. A vast bibliography documents its applications in the food industry.
11. It has been frequently used in the Health industry to produce antigens, toxins, etc...
12. Its simple and inhibitor-free composition makes it suitable for the detection of antimicrobial agents in food and other products.
13. A balanced and high nutrient value together with a lack of fermentable carbohydrates make this medium ideal for maintaining bacterial strains.
14. Classical media for microbiological examination of non-sterile products according to Pharmacopeial Harmonised Methods.

Technique:

Melt the medium contained in the bottles in a water bath or in a microwave oven, avoiding overheating before pouring into Petri dishes when cooled to room temperature.

Once solidified on a flat surface, Spread the plates by streaking methodology or by spiral method.

The inoculated plates are incubated at 30-35 ° C for 24-72 h (bacteria) and 3-5 days for fungi (yeast & molds). Examined daily. (Incubation times greater than those mentioned above or different incubation temperatures may be required depending on the sample, on the specifications,... This medium can be inoculated directly or after enrichment broth).

After incubation, enumerate all the colonies that have appeared onto the surface of the agar.

Each laboratory must evaluate the results according to their specifications.

Calculate total microbial count per ml of sample by multiplying the average number of colonies per plate by the inverse dilution factor if streaked a diluted sample. Report results as Colony Forming Unit (CFU's) per ml or g along with incubation time and temperature.

Quality control**Physical/Chemical control**

Color : yellow

pH: 7.3 ± 0.2 at 25°C

Microbiological control

Melt Medium - Prepare Plates - According to harmonized pharmacopoeial monographs, ISO standards and test methods

Spiral Spreading: Practical range 50 - 100 CFU (Productivity).

Aerobiosis. Incubation at 30-35 °C. Read after 18-24h to 72 h for bacteria and 3-5 days for fungi.

Microorganism**Growth***Escherichia coli* ATCC® 8739, WDCM 00012

Good (≥70 %)

Candida albicans ATCC® 10231, WDCM 00054

Good (≥70 %)

Ps. aeruginosa ATCC® 9027, WDCM 00026

Good (≥70 %)

Staphylococcus aureus ATCC® 6538, WDCM 00032

Good (≥70 %)

Aspergillus brasiliensis ATCC® 16404, WDCM 00053

Good (≥70 %)

Bacillus subtilis ATCC® 6633, WDCM 00003

Good (≥70 %)

L. monocytogenes ATCC® 13932, WDCM 00021

Good (≥70 %)

Salmonella typhimurium ATCC® 14028, WDCM 00031

Good (≥70 %)

Bacillus cereus ATCC® 11778, WDCM 00001

Good (≥70 %)

Sterility Control

Incubation 48 hours at 30-35°C and 48 hours at 20-25°C: NO GROWTH

Check at 7 days after incubation in same conditions

Bibliography

- ATLAS, R.M. & L.C. PARKS (1993) Handbook of Microbiological Media. CRC Press, Inc. London.
- COLIPA (1997) Guidelines on Microbial Quality Management (MQM). Brussels.
- DOWNES, F.P. & K. ITO (2001) Compendium of Methods for the Microbiological Examination of Food, 4th ed, ASM, Washington D.C.
- EUROPEAN PHARMACOPOEIA 8.0 (2014) 8th ed. § 2.6.13. Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. EDQM. Council of Europe. Strasbourg.
- FDA (Food and Drug Administrations) (1998) Bacteriological Analytical Manual. 8th ed. Revision A. AOAC International. Gaithersburg. MD.
- HORWITZ, W. (2000) Official Methods of Analysis of AOAC INTERNATIONAL, 17th ed. Gaithersburg, MD. USA.
- ISO 9308-1 Standard (2000) Water Quality. Detection and enumeration of *E. coli* and coliform bacteria. Membrane filtration method.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 22717 Standard (2015) Cosmetics - Microbiology - Detection of *Pseudomonas aeruginosa*.
- ISO/TS 22964 (2006) Milk and milk products.- Detection of *Enterobacter sakazakii*.
- PASCUAL ANDERSON, M^oR^o (1992) Microbiología Alimentaria. Díaz de Santos S.A., Madrid.
- USP 33 - NF 28 (2011) <62> Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. USP Corp. Inc. Rockville. MD. USA.