

Specification

General purpose solid medium containing animal and plant peptone according to Pharmacopoeial Harmonised Method and ISO Standards.

Presentation

	Packaging Details	Shelf Life	Storage
24 Contact Plates/Ird. Contact Plates - Triple Wrapping with: 15 ± 2 ml	6 contact plates blistered (base of aluminum and PVDC) - 2 Cellophane Bags per blister of 6 plates - 4 blisters of 6 plates (24 plates) per box - each blister exhibits the irradiation indicator (8-14KGy)	7 months	2-25°C

Composition

Composition (g/l):

Peptone from casein	15.0
Soya peptone.....	5.0
Sodium chloride.....	5.0
Agar.....	15.0

Description /Technique

Contact plates are used in the microbiological control of disinfection and cleaning of surfaces. It acts simultaneously as a sampler and incubation culture medium without the need for any other intermediate steps.

The plates come in a form appropriate for this function and can be used with different culture media depending on the type of microbe that needs to be controlled. On average the plates provide a contact surface of approximately 25 cm².

To use, remove the cover and gently press the culture medium on the surface to be controlled, ensuring contact between the two surfaces. The Contact plate is removed and covered with the lid to prevent air contamination. It is advisable that the lid is secured with adhesive tape and the bottom labelled with the sampling data (place, date and time).

If the sample surfaces are rough, the contact plates will not make good contact, even when the pressure is increased. In these cases it is advisable to delineate an sample surface area of 25 cm squared and rub this area vigorously with a wet sterile swab and then rub the swab over the Contact plate.

If verifying the effectiveness of a cleaning or disinfection process, contact plates should be used within two hours after the end of the process, ensuring that the sample surface is dry. It is advisable to always include positive controls, sampling the area before disinfection or dirty areas beside the disinfected area.

The technician will determine the frequency of sampling and disinfection according to performance criteria. Apply the agar directly onto surface to be monitored ensuring that the pressure is distributed over the whole plate for 10 seconds. Clean the surface where the sample was collected in order to remove any traces of agar.

The inoculated plates are incubated at 30-35 ° C for 24-72 h (bacteria) and 3-5 days for fungi (yeast & molds). Examined daily.

Note: Contact plates are used for monitoring the microbiological contamination of surface and air inside cleanrooms, isolators, RABS, food industries and hospitals. The double/triple irradiated wrapping ensures that the package itself doesn't contaminate the environment as the first wrapper is removed just before entering the clean area.

Quality control

Physical/Chemical control

Color : Straw-coloured yellow pH: 7.3 ± 0.2 at 25°C

Microbiological control

Growth Promotion Test according to harmonized pharmacopoeial monographs and test methods & ISO 11133:2014

Inoculate: Practical range 100±20 CFU; Min. 50 CFU (Productivity).

Aerobiosis. Incubation at 30-35 °C. Read after 18-24h to 72 h for bacteria and 3-5 days for fungi.

Microorganism

Growth

<i>Escherichia coli</i> ATCC® 8739, WDCM 00012	Good (≥70 %)
<i>Staphylococcus aureus</i> ATCC® 6538, WDCM 00032	Good (≥70 %)
<i>Bacillus subtilis</i> ATCC® 6633, WDCM 00003	Good (≥70 %)
<i>Ps. aeruginosa</i> ATCC® 9027, WDCM 00026	Good (≥70 %)
<i>Candida albicans</i> ATCC® 10231, WDCM 00054	Good (≥70 %)
<i>Aspergillus brasiliensis</i> ATCC® 16404, WDCM 00053	Good (≥70 %)
<i>Salmonella typhimurium</i> ATCC® 14028, WDCM 00031	Good (≥70 %)
<i>L. monocytogenes</i> ATCC® 13932, WDCM 00021	Good (≥70 %)
<i>Bacillus cereus</i> ATCC® 11778, WDCM 00001	Good (≥70 %)

Sterility Control

Incubation 48 hours at 30-35°C and 48 hours at 20-25°C: NO GROWTH

Check at 7 days after incubation in same conditions

Bibliography

- ATLAS, R.M. & L.C. PARKS (1993) Handbook of Microbiological Media. CRC Press, Inc. London.
- COLIPA (1997) Guidelines on Microbial Quality Management (MQM). Brussels.
- DOWNES, F.P. & K. ITO (2001) Compendium of Methods for the Microbiological Examination of Food, 4th ed, ASM, Washington D.C.
- EUROPEAN PHARMACOPOEIA 8.0 (2014) 8th ed. § 2.6.13. Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. EDQM. Council of Europe. Strasbourg.
- FDA (Food and Drug Administrations) (1998) Bacteriological Analytical Manual. 8th ed. Revision A. AOAC International. Gaithersburg. MD.
- HORWITZ, W. (2000) Official Methods of Analysis of AOAC INTERNATIONAL, 17th ed. Gaithersburg, MD. USA.
- ISO 9308-1 Standard (2000) Water Quality. Detection and enumeration of *E. coli* and coliform bacteria. Membrane filtration method.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 22717 Standard (2015) Cosmetics - Microbiology - Detection of *Pseudomonas aeruginosa*.
- ISO/TS 22964 (2006) Milk and milk products.- Detection of *Enterobacter sakazakii*.
- PASCUAL ANDERSON, M^oR^o (1992) Microbiología Alimentaria. Díaz de Santos S.A., Madrid.
- USP 33 - NF 28 (2011) <62> Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. USP Corp. Inc. Rockville. MD. USA.